

**CLIENT NAME: DEFENCE CONSTRUCTION CANADA
14 ALERT BLVD, 8 WING TRENTON
ASTRA, ON K0K3W0
(613) 392-2811**

ATTENTION TO: Chris McRae

PROJECT NO: Alert

AGAT WORK ORDER: 13T748220

WATER ANALYSIS REVIEWED BY: Parvathi Malemath, Data Reviewer

DATE REPORTED: Aug 23, 2013

PAGES (INCLUDING COVER): 4

VERSION*: 2

Should you require any information regarding this analysis please contact your client services representative at (905) 712-5100

***NOTES**

VERSION 2: Revised Sample IDs: Alt-8 changed to Alt-9 as per clients request (September 3rd 2013)

All samples will be disposed of within 30 days following analysis. Please contact the lab if you require additional sample storage time.



AGAT Laboratories

Certificate of Analysis

AGAT WORK ORDER: 13T748220

PROJECT NO: Alert

5835 COOPERS AVENUE
MISSISSAUGA, ONTARIO
CANADA L4Z 1Y2
TEL (905)712-5100
FAX (905)712-5122
<http://www.agatlabs.com>

CLIENT NAME: DEFENCE CONSTRUCTION CANADA

ATTENTION TO: Chris McRae

Lead (Water)

DATE RECEIVED: 2013-08-16

DATE REPORTED: 2013-08-23

		SAMPLE DESCRIPTION:		ALT-9-MID	ALT-9-END
		SAMPLE TYPE:		Water	Water
		DATE SAMPLED:		8/3/2013	8/3/2013
Parameter	Unit	G / S	RDL	4660492	4660500
Lead	µg/L	0.50	<0.50	<0.50	<0.50

Comments: RDL - Reported Detection Limit; G / S - Guideline / Standard

Certified By:

Parvathi Malenath

Quality Assurance

CLIENT NAME: DEFENCE CONSTRUCTION CANADA

AGAT WORK ORDER: 13T748220

PROJECT NO: Alert

ATTENTION TO: Chris McRae

Water Analysis


RPT Date: Aug 23, 2013

RPT Date: Aug 23, 2013			DUPLICATE			Method Blank	REFERENCE MATERIAL		METHOD BLANK SPIKE			MATRIX SPIKE			
PARAMETER	Batch	Sample Id	Dup #1	Dup #2	RPD		Measured Value	Acceptable Limits		Recovery	Acceptable Limits		Recovery	Acceptable Limits	
								Lower	Upper		Lower	Upper		Lower	Upper

Lead (Water)

Lead	1		24.9	25.1	0.8%	< 0.50	101%	90%	110%	96%	90%	110%	94%	70%	130%
------	---	--	------	------	------	--------	------	-----	------	-----	-----	------	-----	-----	------

Certified By:



Method Summary

CLIENT NAME: DEFENCE CONSTRUCTION CANADA

AGAT WORK ORDER: 13T748220

PROJECT NO: Alert

ATTENTION TO: Chris McRae

PARAMETER	AGAT S.O.P	LITERATURE REFERENCE	ANALYTICAL TECHNIQUE
Water Analysis			
Lead	MET-93-6103	EPA SW-846 6020A & 200.8	ICP-MS

